

L

A

M

P

I

R

A

N


Data grafik LJ glukosa bulan agustus 2021

INTERNAL QUALITY CONTROL CHART

INSTITUTION		LABORATORIUM PATOLOGI KLINIK RS UNS						
TEST NAME		GLUKOSA			INSTRUMENT		TMS 30i	
REAGENT		DIALINE			CONTROL NAME		ProLINE 28657	
METHOD								
PERIOD		Agustus-21		UNIT		mg/dl		
				TARGET VALUE		- 2S	TARGET	+ 2S
						78	90,8	105,5

No.	DATE	C.FACTOR	R.BLANK	VALUE	ERROR
1	08/01/21			90,9	
2	08/02/21			86	
3	08/03/21			85,4	
4	08/04/21			90	
5	08/05/21			83	
6	08/06/21			87,1	
7	08/07/21			95,2	
8	08/08/21			87,6	
9	08/09/21			85,9	
10	08/10/21			84,5	
11	08/11/21			81,8	
12	08/12/21			86	
13	08/13/21			97,5	
14	08/14/21			92,9	
15	08/15/21			89,1	
16	08/16/21			85,5	
17	08/17/21			88,6	
18	08/18/21			85,1	
19	08/19/21			89	
20	08/20/21			80,2	
21	08/21/21			95,2	
22	08/22/21			90	
23	08/23/21			84,7	
24	08/24/21			87	
25	08/25/21			86,4	
26	08/26/21			86	
27	08/27/21			86,4	
28	08/28/21			94	
29	08/29/21			94,4	
30	08/30/21			81,6	
31	08/31/21			88,2	

AVR			87,91
SD			4,27
CV %			4,86



Ver. 1.2 August 2001. Author: Alexander D. Alendo




Grafik LJ kreatinin bulan agustus 2021

INTERNAL QUALITY CONTROL CHART

INSTITUTION	LABORATORIUM RSUP SURAKARTA			INSTRUMENT	TMS 30i		
TEST NAME	KREATININ			CONTROL NAME	ProLINE 28657		
REAGENT	DIALINE			TARGET VALUE	- 2S	TARGET	+ 2S
METHOD					0,8	0,97	1,14
PERIOD	Agustus-21	UNIT	SEL/ul				

No.	DATE	C.FACTOR	R.BLANK	VALUE	ERROR
1	08/01/21			1,04	
2	08/02/21			0,93	
3	08/03/21			0,99	
4	08/04/21			0,92	
5	08/05/21			1,02	
6	08/06/21			1	
7	08/07/21			0,93	
8	08/08/21			0,9	
9	08/09/21			0,98	
10	08/10/21			1,01	
11	08/11/21			1,97	12S 13S R4S
12	08/12/21			0,99	R4S
13	08/13/21			0,87	
14	08/14/21			0,96	
15	08/15/21			1	
16	08/16/21			0,97	
17	08/17/21			0,99	
18	08/18/21			0,87	
19	08/19/21			0,88	
20	08/20/21			0,98	
21	08/21/21			1,06	
22	08/22/21			0,99	
23	08/23/21			1,01	
24	08/24/21			1,07	
25	08/25/21			0,91	
26	08/26/21			0,98	
27	08/27/21			1,08	
28	08/28/21			0,98	
29	08/29/21			1,09	
30	08/30/21			1,03	
31	08/31/21			1,02	


AVR		1,01	
SD		0,19	
CV %		18,45	




Ver 1.2, August 2001, Author: Alexander D. Alvando




Grafik LJ glukosa bulan September 2021




Grafik LJ kreatinin bulan September 2021




Grafik LJ glukosa bulan oktober 2021




Grafik LJ kreatinin bulan oktober 2021




Grafik LJ glukosa bulan november 2021




Grafik LJ kreatini bulan November 2021



Grafik LJ glukosa bulan desember 2021




Grafik LJ kreatinin bulan desember 2021



**DATA QC ALAT TMS
GLUKOSA CREATININE SEPTEMBER 2021**

Tanggal	Glukosa	Creatinine
1	89,0	1,04
2	81,3	0,93
3	90,3	0,99
4	90,2	0,92
5	81,2	1,02
6	90,9	1,00
7	91,9	0,93
8	89,5	0,90
9	91,9	0,98
10	91,2	1,01
11	90,6	0,97
12	91,6	0,99
13	91,9	0,87
14	93,4	0,96
15	90,6	1,00
16	86,7	0,97
17	89,6	0,99
18	91,2	0,87
19	92,1	0,88
20	89,3	0,98
21	87,1	1,06
22	79,5	0,99
23	80,0	1,01
24	89,5	1,07
25	94,6	0,91
26	79,1	0,98
27	90,6	1,08
28	85,8	0,98
29	89,4	1,09
30	91,4	1,03



Range :
Glukosa 78 - 105,5
Creatinine 0,8 - 1,14

03 AUG 2022

**DATA QC ALAT TMS
GLUKOSA CREATININE AGUSTUS 2021**

Tanggal	Glukosa	Creatinine
1	90,9	1,04
2	86,0	0,93
3	85,4	0,99
4	90,0	0,92
5	83,0	1,02
6	87,1	1,00
7	95,2	0,93
8	87,6	0,90
9	85,9	0,98
10	84,5	1,01
11	81,8	0,97
12	86,0	0,99
13	97,5	0,87
14	92,9	0,96
15	89,1	1,00
16	85,5	0,97
17	88,6	0,99
18	85,1	0,87
19	89,0	0,88
20	80,2	0,98
21	95,2	1,06
22	90,0	0,99
23	84,7	1,01
24	87,0	1,07
25	86,4	0,91
26	86,0	0,98
27	86,4	1,08
28	94,0	0,98
29	94,4	1,09
30	81,6	1,03
31	88,2	1,02

Range :

Glukosa 78 - 105,5

Creatinine 0,8 - 1,14

03 AUG 2022



**DATA QC ALAT TMS
GLUKOSA CREATININE NOVEMBER 2021**

Tanggal	Glukosa	Creatinine
1	94,6	1,00
2	92,6	0,83
3	94,0	0,89
4	94,8	0,95
5	92,5	0,92
6	93,0	0,85
7	95,4	0,89
8	101,5	0,91
9	91,2	0,86
10	92,2	0,84
11	95,0	0,87
12	92,5	0,98
13	94,9	0,91
14	91,2	0,98
15	96,2	0,89
16	93,3	1,02
17	99,0	0,99
18	89,6	0,93
19	92,1	0,93
20	88,6	0,94
21	101,3	0,93
22	89,6	0,84
23	92,1	0,85
24	88,6	0,89
25	91,3	0,89
26	93,4	0,90
27	93,9	0,90
28	93,6	0,98
29	97,2	0,89
30	95,1	0,91



Range :
Glukosa 78 - 105,5
Creatinine 0,8 - 1,14

03 AUG 2022

DATA QC ALAT TMS
GLUKOSA CREATININE DESEMBER 2021

Tanggal	Glukosa	Creatinine
1	78,0	1,27
2	79,0	1,28
3	87,0	1,44
4	77,0	1,31
5	76,0	1,28
6	78,0	1,30
7	78,0	1,30
8	77,0	1,28
9	77,0	1,31
10	84,0	1,41
11	79,0	1,30
12	76,0	1,24
13	75,0	1,22
14	81,0	1,30
15	76,0	1,35
16	78,0	1,29
17	85,0	1,44
18	79,0	1,29
19	85,0	1,26
20	77,0	1,29
21	84,0	1,28
22	85,0	1,30
23	82,0	1,29
24	80,0	1,37
25	82,0	1,33
26	82,0	1,46
27	79,0	1,41
28	78,0	1,36
29	85,0	1,39
30	75,0	1,35
31	76,0	1,32



Range :
 Glukosa 70 - 96,7
 Creatinine 1,06 - 1,66

03 AUG 2022

**DATA QC ALAT TMS
GLUKOSA CREATININE OKTOBER 2021**

Tanggal	Glukosa	Creatinine
1	81,6	1,01
2	81,0	1,02
3	89,3	1,00
4	89,2	1,10
5	83,5	1,05
6	85,2	0,98
7	91,0	1,08
8	82,2	1,04
9	80,9	1,01
10	93,6	0,89
11	80,7	0,92
12	81,5	1,02
13	87,9	1,03
14	88,9	1,05
15	82,4	0,94
16	90,2	1,07
17	92,1	1,07
18	80,6	0,96
19	87,6	0,93
20	89,1	0,96
21	86,4	0,97
22	85,6	0,80
23	89,2	0,87
24	90,1	0,90
25	83,8	0,89
26	86,3	0,89
27	91,1	0,78
28	96,1	0,91
29	91,9	0,87
30	95,0	0,89
31	86,6	0,84



Range :
Glukosa 78 - 105,5
Creatinine 0,8 - 1,14

03 AUG 2022



